

DECLARATION of Conformity

Order No.: MT-S0511342

Holder: **Shanghai MCP Corp.**
Rm 908 No. 295 Cao An Road, Shanghai China

Equipment Description: Oscilloscope

Model Name: CDQ620, CQ620R, CQ620B, CQ620BF, CQ620,
CQ620F, CQ620C, CQ620CF, CQ630R, CQ630B,
CQ630BF, CQ630, CQ630F, CQ630C, CQ630CF

Technical Data: AC 110/220V \pm 10, 50/60Hz, 35W(40VA)

The submitted products have been tested by us with the listed standards and found in compliance with the following European Directives:

The LVD directive 73/23/EEC as last amended by EEC directive 93/68/EEC

EN61010-1: 2001
EN61010-031: 2002

The tests were performed in normal operation mode. The test results apply only to the particular sample tested and to the specific tests carried out.

This certificate applies specifically to the sample investigated in our test reference number only.

The CE markings as shown below can be affixed on the product after preparation of necessary technical documentation.

Other relevant Directives have to be observed.

November 25, 2005



Mark Wen
For Chief Executives



Most Compliance Laboratory Limited
21A-B 4th Building, Time City, Chuangye Road, Nanshan
Shenzhen Guangdong, China
Tel: 86 755 8617 0306

Fax: 86 755 8617 0310

