



STATEMENT OF ATTESTATION

No. 304/97

Holder : Precision Mastech Enterprises Co.
 Room 1708-9, Hewlett Centre,
 52 Hoi Yuen Road, Kwun Tong,
 Kowloon, Hong Kong.

Type of Product : DC Capacitor Meter

Model reference : "MASTECH" MY6013A

Electrical rating : 9V DC, 1 x Neda 1604 or 6F22 006p battery

Manufacturer : Shenzhen Huayi Mastech Co. Ltd.
 8/F., East Wing, Blk 4, Saige Science & Technology
 Industrial Garden, Hua Qiang Bei Road,
 Shenzhen China.

Test Report No. : EMC103/97

The above mentioned products fulfil the following standard(s) of EMC Directive 89/336/EEC.

EN 50081-1:1992 in conjunction w/ EN 55022:1994+ A1:1995	Generic emission standard
EN 50082-1:1992 in conjunction w/ IEC 801-2:1991, IEC 801-3: 1984 IEC 801-4:1988	Generic immunity standard

This statement only relates to the sample(s) submitted for testing, and based on the technical documentation and report(s). This statement does not allow the use of any RWTUV test mark on the named product(s).

Signature: Nicholas for RWTUV-HK

Signature: [Signature] for RWTUV-Anlagentechnik

Place / date : Hong Kong, 12 September 1997

