

1997-11-11

**TEST REPORT
NO.: EMC104/97****APPLICANT:**

Precision Mastech Enterprises Co.
Room 1708-9, Hewlett Centre,
52 Hoi Yuen Road, Kwun Tong,
Kowloon, Hong Kong.

DESCRIPTION OF SAMPLE(S):

Product: LC Meter
Manufacturer: Shenzhen Huayi Mastech Co. Ltd.
8/F., East Wing, Blk 4, Saige Science & Technology
Industrial Garden, Hua Qiang Bei Road, Shenzhen China.
Model Number: MY6243
Brand Name: MASTECH
Rating: 9V DC, 1 x Neda 1604 or 6F22 006p battery
Origin: China

INVESTIGATIONS REQUESTED:

Perform EMC tests

RESULT/REMARK:

Please see attached sheet(s).

At the request of the applicant, the EMC tests were performed with the product in capacitance mode.

The EMC tests were performed with the product measuring a 10 μ f electrolytic capacitor. The measurement range was 20 μ f. The 4mm 'banana' sockets were used for the measurement of the capacitor together with the 0.2m long measurement leads supplied with the product.

TEST EQUIPMENT AUDIT AND PHOTO(s):

Please see Appendix A.



Nicholas Yiu
Nicholas Yiu
for RWTUV HK Ltd.

Remark: This test report is transferred result from the work data report no.HM479/503 and ID:HK01